



PEWA
Messtechnik GmbH

Weidenweg 21
58239 Schwerte

Telefon: +49 (0) 2304-96109-0
Telefax: +49 (0) 2304-96109-88
eMail: info@pewa.de
Homepage: www.pewa.de

FLUKE

700BCW Barcode-Lesestift



Description: This Hewlett Packard HBSW-8200 Bar Code Wand reads eight common bar code symbologies. Code 39, Interleave 2 or 5, UPC A and E, EAN 8 and 13, Codabar, Code 128, Code 11, Code 93, and MSI Code. Programmable via bar code labels or escape sequences. LED indicates when a bar code has been correctly decoded. Resolution, 0.19 mm. Usable from -20 °C to 65 °C, in ambient light to 100,00 lux. EMI compliance per FCC part 15 Class B. ESD immunity, 25 kV.

Application: Supports the scanned entry of procedure identification, tag ID, serial number, and operator ID to a Fluke 743 Documenting Process Calibrator.

Included: Bar code wand with integrated decoding circuitry, holder, and users manual; a Fluke 700BCA Bar Code Adapter, and instruction sheet in English, French, German, Italian, and Spanish.

Mechanical & General Specifications

Size	160 x 23 mm. Coiled cable extends to approximately 2.5 m.
Weight	160 g
Warranty	One year
Material	Polycarbonate wand, sapphire tip.