


PEWA Messtechnik GmbH

Weidenweg 21 58239 Schwerte

Tel.: 02304-96109-0 Fax: 02304-96109-88 E-Mail: info@pewa.de Homepage : www.pewa .de

Model 6411 Back Probe Pin

CE



FEATURES:

- Make connections to terminals inside connectors up to 1" (24,39mm) deep.
- \checkmark Pin can be bent to fit into tight spaces
- ✓ Banana Jack fits modular test leads, such as Pomona models 5907A, 5908A and 5909A
- \checkmark Bright yellow color is easy to find in poor light

MATERIALS:

Jack: Brass Tip Insulation: Polypropylene, Yellow Tip: Nickel Plated Beryllium Copper Boot Insulation: Vinyl, Yellow

RATINGS:


For CE compliance: not intended for hand-held use at voltages above 33Vrms / 70Vdc. Current: 3 Amperes Max.

ORDERING INFORMATION:

Model 6411 (10 per package)

All dimensions are in inches. Tolerances (except noted): $.xx = \pm.02$ " (,51 mm), $.xxx = \pm.005$ " (,127 mm). All specifications are to the latest revisions. Specifications are subject to change without notice. Registered trademarks are the property of their respective companies.

Sales: 800-490-2361 Fax: 888-403-3360 Technical Support: 800-241-2060 (technical support@pomonatest.com) For "Where to Buy" information, visit the Pomona web site at www.pomonaelectronics.com





Model 6411 Back Probe Pin

Pomona®


The items included with this product are certified to meet IEC 1010-2-031 safety requirements.

Les élements inclus avec ce produit sont certifies pour répondre à des exigences de surete du IEC 1010-2-031.

Los items incluidos con este producto se certifican para resolver requisitos de seguridad del IEC 1010-2-031.

 Only IEC 1010 type shrouded test plugs should be used with IEC 1010 type jacks. NEVER INSERT A PROBE
TIP WITH BANANA SPRING INTO A TEST ACCESSORY JACK
THAT WILL BE USED AS A HAND HELD PROBING DEVICE -Seuls les vérificateurs de tension renforcés de type IEC 1010 devraient être utilisés avec des jacks de type IEC 1010. NE JAMAIS INSERER DE BOUT DE SONDE AVEC RESSORT BANANE DANS UN JACK D'ESSAI D'ACCESSOIRES QUI SERA UTILISE COMME DISPOSITIF DE SONDAGE A MAIN.

-Se deben usar sólo tapones de prueba reforzados del tipo IEC 1010 con receptáculos de tipo IEC 1010. NUNCA INTRODUZCA UNA CLAVIJA CON UN MUELLE TIPO BANANA EN UN RECEPTACULO DE PRUEBA QUE A SU VEZ SE USE COMO APARATO DE PRUEBA MANUAL.



All dimensions are in inches. Tolerances (except noted): $.xx = \pm .02$ " (,51 mm), $.xxx = \pm .005$ " (,127 mm). All specifications are to the latest revisions. Specifications are subject to change without notice. Registered trademarks are the property of their respective companies.

Sales: 800-490-2361 Fax: 888-403-3360 Technical Support: 800-241-2060 (technical support@pomonatest.com) For "Where to Buy" information, visit the Pomona web site at www.pomonaelectronics.com