



PEWA
Messtechnik GmbH

Weidenweg 21
58239 Schwerte

Tel.: 02304-96109-0
Fax: 02304-96109-88
E-Mail: info@pewa.de
Homepage : www.pewa.de



HI 7634-00 Conductivity Probe

HI 7634-00 is an EC/TDS probe with built-in NTC sensors which provide temperature compensated conductivity measurements.

It can be used with the following Hanna meters:

BL 983313: 0 to 1999 $\mu\text{S}/\text{cm}$

BL 983315: 0.0 to 199.9 ppm

BL 983319: 0 to 1999 ppm

BL 983320: 0.0 to 199.9 $\mu\text{S}/\text{cm}$

BL 983321: 0.00 to 19.99 ppm

BL 983322: 0.00 to 19.99 $\mu\text{S}/\text{cm}$

BL 983324: 0.0 to 49.9 ppm

BL 983329: 0 to 999 ppm

WARRANTY:

HI 7634-00 probe is warranted for a period of one year. This warranty is limited to repair or replacement free of charge.


Damages due to accident, misuse, tampering or lack of prescribed maintenance are not covered.

If service is required, contact the dealer from whom you purchased the instrument. If under warranty, report the model number, date of purchase, serial number and the nature of the failure. If the repair is not covered by the warranty, you will be notified of the charges incurred. If the instrument is to be returned to Hanna Instruments, first obtain a Returned Goods Authorization Number from the Customer Service department and then send it with shipment costs prepaid. If the repair is not covered by the warranty, you will be notified of the charge for repair or replacement. When shipping any instrument, make sure it is properly packaged for complete protection.

www.hannainst.com

CONNECTIONS:

The wires from the probe cable must be connected to the process EC or TDS controller as shown below. The connections are color coded for an easy installation.



HI 7634-00

Note: once the probe is attached to the meter, it is recommended to calibrate it before taking any measurement. Follow the calibration procedure in the meter's manual.

SPECIFICATIONS (typical @25 °C):

Cell Constant	1 cm ⁻¹
NTC sensors	10 KΩ
Materials	PP and AISI316
Temperature	-5 to 50 °C
Max. pressure	3 bar

PROBE MAINTENANCE:

To improve the probe performance and prolong its life, it is recommended to clean it regularly.

- Immerse the tip of the probe in **HI7061** Cleaning Solution for one hour.
- If a more thorough cleaning is required, brush the metal pins with very fine sandpaper.
- After cleaning, rinse the probe with tap water and recalibrate the meter.
- When not in use, clean the probe before storing it away.