

# Oscilloscope Firmware Installation Instruction

## for CombiScope Type HM2008

Together with this information you receive one firmware update for the HAMEG oscilloscope mentioned above. The aim of the firmware update is that the oscilloscope reaches the latest firmware revision, with the highest version number.

For updating a PC with Microsoft Windows 2000 or XP, a free COM port or USB port is required.

### **1. Firmware version**

The actual firmware version already on the scope is displayed after switching on if „Quick Start“ is off. The „Quick Start“ function can be changed after pressing the SETTINGS pushbutton and calling „Misc“.

The update name consists of oscilloscope type and version (e.g. HM2008\_05.303-yy.yyy.exe). The following „yy.yyy“ is not of interest for the update.

Where errors would occur by trying to update an older version, this action will be automatically prevented.

## 2. Hardware

PC and oscilloscope must be connected via a screened cable (e.g. HZ14) with a 9pole Sub-D male connector on one side and a 9pole Sub-D female connector on the other side. All pins must be connected 1:1 (pin1 with pin 1, pin 2 with pin 2 etc.). Connect the 9pole Sub-D female connector with a free COM port of the PC and the 9pole Sub-D male connector with the RS-232 connector at the rear side of the oscilloscope. Alternatively a USB cable can be used, if at the PC and oscilloscope a USB port is available.


## 3. Firmware Installation

The update program is a simple executable program that must not be installed.


### Attention!

It is possible that an automatical adjustment will be made during the update process. Thus the oscilloscope must have reached its normal operating temperature after 20 minutes warm up and no signal must be applied at the oscilloscope inputs.


1. Start the update program for your oscilloscope under consideration of item 1 (update selection) by opening the directory Firmware and clicking the right instrument type and version. This opens a window.




2. Check the interface settings by clicking „Settings“, select „Serial Ports“ and click „Options“.




3. This opens a second window that shows the baud rate. If 115200 baud is displayed click „OK“ to confirm and close that window.




4. Click „Settings” once again, select „Serial Ports” and click „Activate”.




5. Thereafter the update software searches for HAMEG oscilloscopes and displays the reply in the window.




6. Click on the line in which the oscilloscope to be updated is displayed. That marks the line and additionally the button „Start Update“ is displayed. Click that button to start the update.



7. This opens a second window that shows the update progress and thereafter other information. When flash memory update is displayed the oscilloscope must not be switched off.




It is also possible that the following window will be displayed. If the required conditions are met, the firmware update can be continued with „Yes“.



After the firmware has been updated an auto restart is made and the update is checked.

8. A green tick briefly displayed in front of the oscilloscope type indicates the end of the update procedure.



HAMEG Instruments GmbH  
Mainhausen, September 2007

Oscilloscopes



Spectrum Analyzers



Power Supplies



Modularsystem  
Series 8000



Programmable Instruments  
Series 8100



authorized dealer

[www.hameg.com](http://www.hameg.com)

Subject to change without notice  
13-09-2007-gw

© HAMEG Instruments GmbH  
A Rohde & Schwarz company  
® registered trademark



DQS-Certification: DIN EN ISO 9001:2000  
Reg.-Nr.: DE-071040 QM

HAMEG Instruments GmbH  
Industriestraße 6  
D-63533 Mainhausen  
Tel +49 (0) 61 82 800-0  
Fax +49 (0) 61 82 800-100  
[sales@hameg.de](mailto:sales@hameg.de)