

INSTRUCTION MANUAL

TT-SI 9010

70MHz Active Differential Probe



PEWA
Messtechnik GmbH
Weidenweg 21
58239 Schwerte
Tel.: 02304-96109-0
Fax: 02304-96109-88
E-Mail: info@pewa.de
Homepage: www.pewa.de

TESTEC

These probe is in compliance with IEC-61010-031 CAT I, Pollution Degree 2

1. Safety Terms and Symbols

Terms appear in this manual:



WARNING. Warning statements identify conditions or practice that could result in injury or loss life.



CAUTION. Caution statements identify conditions or practice that could result in damage to this product or other property.

Safety Symbols



Connect it to safety earth ground using the wire recommended in the user's manual.



High voltage danger



The symbol on an instrument indicates that the user should refer to the operating instructions located in the manual.

2. General Safety Summary

Review the following safety precautions to avoid injury and prevent damage to this probe or any products that connected to it.

Observe Maximum Working Voltage

To avoid any injury, do not use the probe under the condition that the voltage between either input lead or earth is above 2500Vrms CAT I. This voltage rating applies to both settings 1/100 & 1/1000.

Must be Grounded

This probe is grounded with the shell of BNC connector and an auxiliary grounding terminal, through the grounding conductor of the power cord of the measurement instrument. Before making connections to the input leads of this probe, ensure that the output BNC connector is attached to the BNC connector of the measurement instrument and the auxiliary grounding terminal is connected to a proper ground, while the measurement instrument is properly grounded.

Do Not Operate Without Covers

To avoid electric shock or fire hazard, do not operate this probe with covers removed.

Do Not Operate in Wet/Damp Conditions

To avoid electric shock, do not operate this probe in wet or damp conditions.

Do Not Operate in Explosive Atmosphere

To avoid injury or fire hazard, do not operate this probe in an explosive atmosphere.

Avoid Exposed Circuit

To avoid injury, remove jewelry such as rings, watches, and other metallic objects. Do not touch exposed connections and components when power is present.

Use Proper Power Source

To ensure this probe functions well, use four AA cells or 6VDC/60mA or regulated 9VDC/40mA mains adaptor or power lead. Do not operate this probe from a power source that applies more than the voltage specified.

Do Not Operate With Suspected Failures

If you suspect there is damage to this probe, have it inspected by qualified service personnel.

3. Description

By enabling conventional oscilloscopes to display and measure in-circuit waveforms that are referenced to high common mode voltages. The differential probe extends the measurement capability of oscilloscopes in electronic power converters, inverters, motor speed controls, switch mode power supplies, and many applications.

4. Installation

- a. Simply plug-in the BNC output connector to the vertical input of a general purposed oscilloscope or other measurement instrument, and connects the auxiliary grounding terminal to a proper ground. The measurement instrument must have a ground referenced.
- b. Connect an appropriate power source to this probe and or enter the batteries, then turn it on.
- c. Select the proper attenuation ratio. When measuring signals below 700V switch the attenuation ration to 1/100 in order to get higher resolution and less noise ratio. Otherwise, set the attenuation ratio to 1/1000.



WARNING. To protect against electric shock, use only the accessories supplied with this probe.

- d. Using the appropriate probe accessories, connect the inputs to the circuits under measurement.



CAUTION. This probe is to carry out differential measurement between two points on the circuit under measurement. This probe is not for electrically insulating the circuit under measurement and the measuring instrument.

5. Appearance


The differential probe looks as follows.



- a. Output Cable The BNC output connector and an auxiliary grounding terminal are connected to the oscilloscope.
- b. Input Leads The input leads of the differential probe connect to the HV alligator clips that come with the probe.
- c. HV Alligator Clips For safe test point connection

6. Available Power Sources

- a. 4 x AA batteries
- b. Mains adaptor (6VDC/60mA or regulated 9VDC/40mA),
- c. Lemo® Power Cord, for oscilloscopes with power output - Lemo® connector.
- d. Probus® Power Cord, for oscilloscopes with power output - Probus® connector.
- e. USB Power Cord, for oscilloscopes which offer USB connector.



7. Accessories

| Type | Order-No. | Description |
|--------------|-----------|----------------------------------|
| TT-SI NT | 15100 | Mains Adapter |
| TT-SI HC | 15160 | Hardcase |
| TT-SI PROBUS | 15150 | Power Lead with PROBUS-Connector |
| TT-SI LEMO | 15151 | Power Lead with LEMO-Connector |
| TT-SI USB | 15152 | Power Lead with USB-Connector |
| TT-SI EPL1 | 15140 | 1 to 3 Power Splitter |
| TT-SI EPL2 | 15141 | 1 to 4 Power Splitter |


8. Specifications

| | TT-SI 9010 |
|---|---|
| Bandwidth | DC to 70MHz (-3dB) |
| Attenuation Ratio | 1:100 / 1:1000 |
| Accuracy | ±2% |
| Rise Time | 5ns |
| Input Impedance | 10MΩ // 10pF each side to ground |
| Input Voltage - Differential Range | 1:100 ±700V (DC+peak AC) or 700Vrms 1:1000 ±7000V (DC+peak AC) or 5000Vrms |
| Input Voltage - Common Mode Range | 1:100 and 1:1000 ±7000V (DC+peak AC) or 2500Vrms |
| Input Voltage - Absolute Max. Common Mode | 1:100 and 1:1000 ±7000V (DC+peak AC) or 2500Vrms |
| Input Voltage - Absolute Max. Differential Mode | 1:100 and 1:1000 ±7000V (DC+peak AC) or 5000Vrms |
| Measurement Category | CAT I |
| Output Voltage - Swing | ±7V (into 50kΩ load) |
| Output Voltage - Offset (typical) | <±5mV |
| Output Voltage - Noise (typical) | 0,9mVrms |
| Source Impedance (typical) | 50Ω (for using 1MΩ input system oscilloscope) |
| CMRR (typical) | -80dB @50Hz, -60dB @20kHz |
| Ambient Operating Temperature | -10°C to 40°C |
| Ambient Storage Temperature | -30°C to 70°C |
| Ambient Operating Humidity | 25% to 85% RH |
| Ambient Storage Humidity | 25% to 85% RH |
| Power Requirements - Standard | 4 x AA Cells |
| Power Requirements - Optional | Power lead or Mains Adapter (6VDC/60mA or regulated 9VDC/40mA) |
| Length of BNC Cable | 90cm |
| Length of Input Leads | 60cm |
| Weight | 500g |
| Dimensions (LxWxH) | 202mm x 83mm x 38mm |

- a. The supplied voltage must be less than 12V and greater than 4.4V, otherwise the probe could be damaged or can't be operated properly.
- b. Polarity is "+" inside and "-" outside. For wrong polarity, built-in circuit protects the probe, no danger or damage will occur.
- c. When the voltage of the cells become too low, the power indicator on the will flicker.

9. Derating Curve

The derating curve of the absolute maximum input voltage in common mode is shown as follows.



10. Offset Adjustment

If the offset voltage is too large, short the input leads and turn the adjust variable resistor (DC voltage adjustment) which you find in the hole of the panel by using a flat-head screwdriver until the offset voltage is lowest.

11. Inspection Procedure

- Connect the BNC output connector to the vertical input of a general purposed oscilloscope.
- Install four AA cells or connect an appropriate mains adaptor or power lead to the correct line voltage.
- Set the oscilloscope input coupling to DC and the 1V/div. Center the trace on the display.
- Connect the inputs of the probe to power lines.
- Set the range of the probe to 1/100.
- Then, a 50Hz/60Hz sine-wave of proper amplitude will be displayed on the screen of the oscilloscope and this means the probe is working properly.

12. Cleaning

Use a soft cloth to clean the dirt. Prevent damage to probe.

- Avoid immersing the probe.
- Avoid using abrasive cleaners.
- Avoid using chemicals contains benzene or similar solvents.